

CE marking	Class I medical device in accordance with EU 2017/745		CE MD CLASS I
CH REP	Swiss AR Services AG, Industriestrasse 47, 6300 Zug, Switzerland		
EN	Medical face masks – requirements and test methods		Type IIR
	Test	Result	
	Bacterial filtration efficiency (BFE), (%)	> 99%	
	Pressure difference (Pa/cm2)	< 60	
	Microbiological purity (KBE/g)	≤ 30	
VO (EU) 2023/988	Regulation (EU) Nr. 2023/988 on general product safety.		
VO (EG) 1935/2004	Regulation regarding plastic materials and articles intended to come into contact with food.		
VO (EU) 10/2011	Regulation on plastic materials and articles intended to come into contact with food		